

L-CHECK-1

Immunochromatographic rapid test for the detection of Luteinizing Hormone (LH) in whole blood, plasma, or serum samples Ref. 5091

FOR EASY READER[®] AND EASY READER+[®] USE ONLY

I- PRINCIPLE

The Luteinizing hormone (LH) is a glycoprotein produced by the anterior pituitary in response to luteinizing releasing hormone (LH-RH) secreted by the hypothalamus. In women the LH and FSH (follicle stimulating hormone) are subject to the complex ovulation cycle. The increase and release of LH appears 12 to 18 hours before ovulation occurs.

In men, LH stimulates the interstitial cells (Leydig cells) to produce testosterone.

L-CHECK-1 is a rapid quantitative assay for the detection of human luteinizing hormone LH in whole blood, plasma or serum samples.

As the test sample flows through the absorbent device, the labelled antibody-dye conjugate binds to the LH binding site, forming an antibody-antigen (LH) complex. This complex in continuation binds to the anti-LH antibody in the positive reaction zone (T) and produces a strong pink-rose colour band. In the absence of LH, there is no band appearing in the positive reaction zone (T).

Unbound conjugate irrespectively of LH concentration binds to the reagents in the control zone (C), producing a strong pink-rose colour band, demonstrating that the reagents are functioning correctly.

II- L-CHECK-1 KIT COMPONENTS

Each kit contains everything needed to perform 10 or 20 tests.

1- L-CHECK-1 reaction devices:	10	20
2- Disposable plastic pipettes:	10	20
3- Diluent in a dropper bottle:	2.5 mL	5 mL
4- Instruction leaflet:	1	1

5- Controls (Optional):

Positive control ref. V7000 and Negative control ref. V7001: a freeze-dried preparation of a non-infectious compound in diluted human serum, tested and found negative for anti-HIV, anti-HCV and HBs antigen, containing 0.05 % sodium azide is optionally available as a positive and negative control (1x 0.25 mL). The concentration range is indicated on the vial label.

III- STORAGE AND STABILITY

1- All L-CHECK-1 kit components, including optional control before reconstitution with distilled water, should be stored at any temperature between +4°C and +30°C in their original package..

2- **Do not freeze the test kit.**

3- L-CHECK-1 is stable until the expiry date stated on the package label.

IV- PRECAUTIONS

- 1- For *in vitro* diagnostic use and for professional use only.
- 2- Read the instruction notice carefully before using the test.
- 3- Handle all specimens as if they contained infectious agents. When the assay procedure is completed, dispose of specimen carefully after autoclaving them for at least one hour. Alternatively, they can be treated with 0.5% to 1% solution of sodium hypochlorite for one hour before disposal.
- 4- Wear protective clothing such as laboratory coats and disposable gloves while assaying samples.
- 5- Do not eat, drink or smoke in the area where specimens and kit reagents are handled.
- 6- Avoid any contact between hands and eyes or nose during specimen collection and testing.
- 7- Do not use the test from a damaged protective wrapper.
- 8- Do not use beyond the expiry date which appears on the package label.

V- SPECIMEN COLLECTION AND PREPARATION

- 1- L-CHECK-1 is to be performed on human serum, plasma or whole blood.
- 2- The specimen should be collected under the standard laboratory conditions (aseptically in such a way as to avoid haemolysis).
- 3- **If anticoagulant is needed, only citrate, EDTA or heparin (ammonium or lithium) should be used.**
- 4- Each specimen should be treated as if potentially infectious.
- 5- **Whole blood samples should be tested immediately (< 4 hours). Finger prick samples should be assayed just after collection.**
- 6- If the test is to be run within 48 hours after collection the specimen should be stored in the refrigerator (+2°C to +8°C). If testing is delayed more than 48 hours, the specimen should be frozen. The frozen specimen must be completely thawed, thoroughly mixed and brought to room temperature prior to testing. Avoid repeated freezing and thawing.
- 7- In case of cloudiness, high viscosity or presence of particulate matter into the serum specimen, it should be diluted with equal volume (V/V) of diluting buffer (not provided but available upon request) before testing.



VI- ASSAY PROCEDURE

a) Controls testing


- Wait for 15 minutes after the freeze-dried control has dissolved.
- Add the requested volume (25µL) with **lab pipette (disposable tips)** into the sample well of the cassette and proceed in the same way as for a patient's sample.
- The expected concentration level (**in mIU/mL**) is indicated on the vial label and obtained result must match the indicated value. The concentration level can change slightly depending on lot number.
- **The reconstituted vial should be kept at +2°C to +8°C and should be used within 7 days after reconstitution.**

b) Samples testing

Follow the below instructions or refer to the picture n°1

- 1– Allow samples and L-CHECK-1 test devices to come to room temperature prior to testing.
- 2– Remove the reaction device from its protective wrapper by tearing along the split.
- 3– Label device with the patient's name or control number.
- 4– Fill the plastic pipette with specimen (serum, plasma or whole blood) and by holding it vertically, dispense one drop (25 µL) into sample well if serum or plasma is used. If whole blood is used, dispense 2 drops (50 µL) into the well (▷) **and wait for the blood sample to be completely absorbed before adding diluent.**
- 5– Add exactly 4 drops of diluent (150 µL) in the sample well (▷).
- 6– Read the result (**in mIU/mL**) after 10 minutes either using the immediate or countdown reading mode (see corresponding leaflet).

For general instructions describing how to use the VEDALAB's readers, refer to the corresponding leaflet.



Picture n°1

VII- PERFORMANCES CHARACTERISTICS

a) Linearity

The measuring range is 5-400 mIU/mL.

For LH concentration below 5 mIU/mL, the result will be given as “< 5 mIU/mL”.

For LH concentration over 400 mIU/mL, the result will be given as “> 400 mIU/mL”.

For samples whose concentration is higher than 400 mIU/mL, dilute with saline and repeat the assay as per instructions of Part VI.

b) Accuracy

A study has been performed using serum samples obtained from dilutions of LH international reference material WHO (80/552) covering a range of 5 to 400 mIU/mL. Optical densities expressed as a function of LH concentrations are described by following linear curve:

$$Y = 15.3311 + 1.8818x - 2.0315 \cdot 10^{-3} x^2 \quad (r = 0.98).$$

The results show a good correlation ($r > 0.98$) of the values obtained with L-CHECK-1 on VEDALAB's readers.

c) Sensitivity

Concentrations close to 3 mIU/mL are detected by L-CHECK-1 test. In these cases, results will be rendered as “< 5 mIU/mL”.

Expected LH concentration in male (adults) and female (follicular and luteal phases) are below 25 mIU/mL. For mid-cycle and postmenopausal females, the LH concentration is in the range of 34-90 mIU/mL and 10-65 mIU/mL respectively.

d) Precision

A correlation study was performed on 26 known serum samples preassayed using the ROCHE COBAS analyser. The overall correlation is 92% between VEDALAB L-CHECK-1 and ROCHE COBAS analyser.

e) Interferences

There was no interference observed for FSH (250 mIU/mL), TSH (100 μ IU/mL) and hCG (10 K mIU/mL) hormones at respective concentrations.

f) Intra-assay reproducibility

Within run reproducibility was evaluated using 25 replicates of three sera containing 25, 100 and 200 mIU/mL of LH. The obtained CVs (coefficient of variation) are 13.4%, 14.2% and 12.5% respectively.

VIII- LIMITATIONS

1- As for any diagnostic procedure, the physician should confirm the data obtained using this test by other clinical methods.

2-Whole blood samples should be tested immediately (< 4 hours). Finger prick samples should be assayed just after collection.

3- Some serum specimens with high rheumatoid factor concentration (RF) or C-reactive protein (CRP) may yield non specific positive results during testing. Such cases should be considered before testing.

4- The test is designed to eliminate the potential interference of human antibodies to murine IgG (HAMA). However high level of HAMA could give falsely positive results.

5- This format of test is to be only used with VEDALAB rapid tests readers (EASY READER[®] or EASY READER+[®]).


6- If the reading time (10 minutes) is not strictly respected, wrong results will be obtained.

7- This format of test should not be used for visual reading.

8- As it is true for any diagnostic method or for any measurements through analysers, there is a variability of the obtained result. Therefore, a confidence range of +/- 25% should be considered for the final value and for the clinical significance of the result.

X- BIBLIOGRAPHY

1. Elkin-Hirsch, K., Goldzieher, J.W., Gibbons, W.E. and Besch, P.K. Obstetrics and Gynecology 67 : 450-453 (1986).
2. Garcia, J.E., Jones, G.S., and Wright, G.L. Fertility and sterility 36 : 308-315 (1981).
3. Khan, S.A., Qazi, M.H., and Diczfalusy, E.J. Endocrinol. Invest. 7 : 1-22 (1984).
4. Singh, M., Saxena, B.B., and Rathman, P. Fertility and sterility 41 : 210-217 (1984).

	Read the instructions before use		For <i>in vitro</i> diagnostic use
	Temperature limitations		Do not reuse
	Manufacturer		



Manufactured by VEDALAB - France