

NEPHSTAR™ Cystatin C (Cys C) Kit

Catalog No.

DK066

1. Intended Use

This product is used on NEPHSTAR™ protein analysis system for quantitative determination of human **Cystatin C (Cys C)** in serum, plasma as an aid in diagnosis and treatment of renal diseases.

2. Summary

Cystatin C is a small, 13kDa, non-glycosylated basic protein belonging to the cystatin super-family of cysteine protease inhibitors. Cystatin C is produced by virtually all nucleated cells, and is present in all investigated body fluids. The production is constant and is unaffected by inflammatory processes, gender, age and muscle mass. It is freely filtered through a normal Glomerular membrane, but is then reabsorbed and almost entirely catabolized in the proximal tubules, no Cystatin C returns to the blood. So, the Cystatin C concentration in human blood is closely related to GFR. Therefore, Cystatin C is considered a useful endogenous marker for the detection of renal impairment, numerous studies have shown that serum Cystatin C is superior to serum creatinine as a marker for GFR.

3. Test Principle

Particle-enhanced immunonephelometry is applied. This method involves measuring the light scattered by insoluble complexes formed by reaction between specific protein in samples and its respective antibody covalently coupled to latex particles, and the amount of scattered light is directly proportional to the concentration of the protein under condition that antiserum is in excess. The latex particles increase the size of complexes formed and thus the amount of light as well as the test sensitivity. Concentrations are automatically calculated by reference to a calibration curve stored in the instrument.

4. Kit Components

Code	Name	Volume/Quantity
DA022	Cys C Antiserum	2×1.0 ml
DB022	Cys C Reaction buffer	1×25.0 mL
DC022	Cys C Magnetic card	1
DM022	Cys C Control	1×0.3mL
	Manual	1

5. Materials required but not supplied

- 5.1 NEPHSTAR Protein analysis system (NS100)
- 5.2 NEPHSTAR Accessory pack (DK110)
- 5.3 Electronic pipette (YB201)
- 5.4 Pipette 5-50uL (YB301)
- 5.5 Pipette 100-1000uL(YB302)
- 5.6 Equipment for collection of Samples

6. Storage and Stability

The unopened reagent kit should be stored under 2-8°C and can be used until the expiry date labeled on the kit. Do not freeze! The buffer should be equilibrated to room temperature before use. Once opened store the antisera and control at 2-8°C and the buffer at **18-25°C and be sure to screw on the cap tightly**. Under these conditions the buffer is stable for 3 months, antisera and control for 1 month.

7. Sample Collection And Preparation

Use serum, heparinized plasma and EDTA plasma samples. samples may be stored at 2-8°C for 48 hours, otherwise freeze at -20°C or below; do not freeze and thaw sera more than once. Sample dilutions should be freshly prepared on the day of assay. Testing of the following types of samples may result in misleading values:

- 7.1 Highly lipemic, turbid and haemolysed samples are not suitable for nephelometric assays and should not be used unless centrifuged or prepared using other methods. If the background is too turbid and can not be removed, please think of other measuring method.
- 7.2 Testing of samples containing rheumatoid factors, paraproteins or circulating immunocomplexes can result in misleading values due to non-specific scattering light possibly generated by these articles.

8. Test Procedure

Summary: Reagent volumes added to the cuvette

Reagent	Volume
Sample (1/5)	30ul
Cys C Reaction Buffer	400ul
Cys C Antiserum	40ul

- 8.1 Switch NEPHSTAR on.
- 8.2 Enter chemistry number. Enter chemistry number of Cys C kit (**Cys C=66**). If Cys C assay has never been performed on the instrument before, please swipe card when "please swipe card" is displayed.
- 8.3 The assay name and lot of reagent are displayed. Check carefully, press ENTER if the lot number is identical to that printed on the card or kit label, otherwise swipe card to update the curve data stored in NEPHSTAR.
- 8.4 Dilute samples or controls using NEPHSTAR Sample Diluent supplied in NEPHSTAR Accessory pack (Cat. No: DK110). The default dilution scheme for Cys C assay is 1/5 (e.g. **200uL sample diluent + 50uL sample**).
- 8.5 Prepare one cuvette for each sample to be assayed. Place a stirring bar to the cuvette using the forceps supplied with NEPHSTAR, then add **30uL** of diluted sample carefully to the bottom of the cuvette.
- 8.6 Enter sample ID. Press number keys to enter the sample ID; or press ENTER to accept the currently displayed sample ID.
- 8.7 Enter sample dilution: **5**. Accept the default sample dilution by pressing ENTER, otherwise press number keys to alter the dilution scheme.
- 8.8 Place cuvette in chamber. Place the cuvette containing a stirring bar and **30uL** of diluted sample in the chamber and press it down slightly until it reaches the bottom of the chamber. The cuvette will be detected automatically.
- 8.9 Add reagent. Add **400 uL** Cys C reaction buffer and **40 uL** Cys C antiserum simultaneously into the cuvette using the electronic pipette (Cat. No.: YB201) supplied with NEPHSTAR. NEPHSTAR will sense the addition of reagents. With movement of the stirring bar, the assay begins after blanking and result will be printed automatically at the end of the assay.
- 8.10 On completion of the assay remove the cuvette, press ENTER to perform the next assay. Sample ID will increase sequentially. For alteration of the ID press BACK twice and tip in the right number.
- 8.11 If NEPHSTAR indicates result is higher than measurement range, reassay the sample at a higher

dilution of e.g. 1/40 (210µL sample diluent + 30µL 1/5diluted sample) . Accordingly the sample dilution should be altered to 1/40 (press BACK and then the number keys to alter the sample dilution) .

- 8.12 If NEPHSTAR indicates result is lower than measurement range, reassay the sample at a lower dilution of e.g. 1/1. Accordingly the sample dilution should be altered to 1/1 (press BACK and then the number keys to alter the sample dilution) .
- 8.13 On completion of all assays of the same chemistry press ESC and return to step 8.2. Enter new chemistry number and begin another assay.

9. Quality Control

In accord with good laboratory practice, users should run control with every batch of samples. Results of control should fall in the validity range labeled on the control vial.

10. Sensitivity and measuring range

The sensitivity limit is 0.4mg/L and the upper limit is 7.5mg/L when the default dilution scheme is applied. The sensitivity limit is 0.08mg/L when samples are diluted at 1/1.

11. Antigen Excess

Sample concentration of less than 35mg/L will not result in antigen excess, when the results will be misleadingly low.

Reference Range

12.1 Normal range of Cys C concentration of healthy adult is : for individuals 1-50years:0.55-1.15mg/L, for individuals >50years:0.63-1.44mg/L. We recommend local reference ranges are produced.

12.2 Diagnosis and treatment can not only depend on determination of Cys C. The clinical symptoms and other laboratory findings of respective patients should be taken into consideration.

12. Precision

Two analyte concentrations are assayed within several days using this kit of the same lot on NEPHSTAR. 20 repeat assays are performed for each concentration. The average coefficient variations (CV) for each concentration are displayed in the following table:

Cys C (mg/L)	CV (%)
1.1	1.35
3.2	3.20

13. Correlation Study

A correlation study is performed on 20 clinical serum samples using this kit on NEPHSTAR and Behring Cys C reagent on BNII. The linear regression equation and correlation equation got as showed below demonstrate a good correlation between the two methods:

$$Y=1.08X-0.23$$

(Y= NEPHSTARTM Cys C, X=BNIIITM Cys C)

Correlation coefficient $r=0.998$

14. Caution And Warning

15.1 The reagents are only for in vitro diagnostic use.

15.2 The reagents can be used only by trained personnel and good laboratory practice (GLP) and the stated procedure should be abided strictly.

15.3 All sera have been tested to be HIV(1&2) antibody negative, HBsAg negative. However, the performed testing

method can not assure the absolute absence of infectious agents in blood products, so please be sure to handle the blood products such as controls and antisera as potentially infectious sources.

15.4 All reagents of the kit contain sodium azide as preservative. Take caution when handling them. Ingest or contact of the reagents with skin or mucous membranes is forbidden. Wash with large amount of water and seek medical advice if contact does occur. In addition, explosive metal azides may be formed with lead or copper plumbings; when disposing the reagents be sure to flush with large amount of water to avoid buildup of azide.

15.5 All components of kit are NEPHSTAR™ specific. Reagents of different lots are not interchangeable, otherwise the results can be misleading.



Goldsite Diagnostics Inc.


#3A & 4A, Technology Building Annex

Nanhai Road, Nanshan District

Shenzhen, P.R.China, 518067

Tel: 86 755 26890807

Fax: 86 755 26890799



CMC MEDICAL DEVICES & DRUGS, S.L.

C/ Horacio Lengo No 18, CP 29006, Málaga-Spain